

Specifications for cables from HV PS to Patch Panel Board (revised)

Channel	Pin at PS	PS to PPB Cable	Pin at PPB
0 HV	2	1 1 2	1
0 RET	13	1 7 1	2
1 HV	3	2 2 2	3
1 RET	14	1 8 1	4
2 HV	4	3 2 2	5
2 RET	15	7 7 7	6
3 HV	5	4 2 2	7
3 RET	16	1 9 1	8
4 HV	6	5 2 2	9
4 RET	17	2 2 2	10
5 HV	7	6 2 2	11
5 RET	18	8 8 8	12
6 HV	8	7 2 2	13
6 RET	19	9 9 9	14
7 HV	9	8 2 2	15
7 RET	20	3 3 3	16
8 HV	50	9 2 2	17
8 RET	39	1 1 1	18
9 HV	49	0 3 3	19
9 RET	38	2 1 1	20
10 HV	48	1 3 3	21
10 RET	37	4 4 4	22
11 HV	47	2 3 3	23
11 RET	36	5 5 5	24
12 HV	46	3 3 3	25
12 RET	35	1 3 3	26
13 HV	45	4 3 3	27
13 RET	34	4 1 1	28
14 HV	44	5 3 3	29
14 RET	33	5 1 1	30
15 HV	43	6 3 3	31
15 RET	32	6 1 1	32
SL 1	22	6 6 6	33
SL 2	30	1 1 1	34

Pin 1 - conductor 20

Pin 11 - conductor 10

Pins 14, 51 - conductor 37

Specifications for cables from HV PS to Patch Panel Board (from manufacturer's specs)

Channel	Pin at PS	PS to PPB Cable	Pin at PPB
0 HV	2	1 2 2 1	1
0 RET	13	1 1 1 1	2
1 HV	3	1 3 3 1	3
1 RET	14	1 2 1 2	4
2 HV	4	1 4 4 1	5
2 RET	15	1 3 1 3	6
3 HV	5	1 5 5 1	7
3 RET	16	1 4 1 4	8
4 HV	6	1 6 6 1	9
4 RET	17	1 5 1 5	10
5 HV	7	1 7 7 1	11
5 RET	18	1 6 1 6	12
6 HV	8	1 8 8 1	13
6 RET	19	1 7 1 7	14
7 HV	9	1 9 9 1	15
7 RET	20	1 8 1 8	16
8 HV	50	1 7 3 3 7 1	17
8 RET	39	1 8 2 2 8 1	18
9 HV	49	1 6 3 3 6 1	19
9 RET	38	1 7 2 2 7 1	20
10 HV	48	1 5 3 3 5 1	21
10 RET	37	1 6 2 2 6 1	22
11 HV	47	1 4 3 3 4 1	23
11 RET	36	1 5 2 2 5 1	24
12 HV	46	1 3 3 3 3 1	25
12 RET	35	1 4 2 2 4 1	26
13 HV	45	1 2 3 3 2 1	27
13 RET	34	1 3 2 2 3 1	28
14 HV	44	1 1 3 3 1 1	29
14 RET	33	1 2 2 2 2 1	30
15 HV	43	1 0 3 3 0 1	31
15 RET	32	1 1 2 2 1 1	32
SL 1	22	1 9 1 9 1 1	33
SL 2	30	1 0 2 2 0 1	34